

CERTIFICATE OF ANALYSIS

PRODUCT PP2010
BISMUTH SULPHITE

LOT NUMBER 4338704

EXPIRY DATE 2021.11.19

RELEASE DATE 2021.11.11

Physical Testing	Results	Specification
------------------	---------	---------------

Appearance (Colour)	Pass	Cream
Clarity	Pass	Opaque
Gel Strength	Pass	Satisfactory
pH	Pass	7.4 - 7.8
Microbial Load @ 30°C for 3 days /sample	Pass	AQL 2.5

Microbiology Testing	Result	Specification
----------------------	--------	---------------


Salmonella typhimurium ATCC® 14028	Pass	Black colony with met sheen
Salmonella subsp. 3b IMVS1679	Pass	Black colony with met sheen
Salmonella hofit IMVS 1799	Pass	Black colony with met sheen
Enterococcus faecalis ATCC® 29212	Pass	Inhibited

Tested in accordance with QSP1105 on a representative sample of the batch. Results obtained at the time of release.

Decision rules are outlined in the product Technical Specification.

Uncertainty of measurement introduced during quantitative testing has been determined. Values not reported on the certificate of analysis, but details available on request.

Released by: Paul Costello



.....
Approved Signatory

Accredited for compliance with ISO/IEC 17025 - Testing.
This report may not be reproduced except in full.



Accreditation
Number 3186

MANUFACTURED AND TESTED BY:
Thermo Fisher Scientific, Oxoid Australia Pty Ltd
20 Dalgleish Street, Thebarton, South Australia 5031
Tel: Toll Free 1300 735 292, Fax: 1800 007 054
Email: microtechsupport@thermofisher.com
Website: www.thermofisher.com