

**CERTIFICATE OF ANALYSIS**

**PRODUCT** PP2249  
BRILLIANCE UTI/HBA

**LOT NUMBER** 4305495

**EXPIRY DATE** 2021.05.07

**RELEASE DATE** 2021.03.25

**Physical Testing Side 1** **Results** **Specification**  
*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Appearance (Colour) - HBA	Pass	Red
Clarity - HBA	Pass	Opaque
Gel Strength	Pass	Satisfactory
pH - HBA	7.4	7.2 - 7.6

Microbiological Testing Side 1	Result	Specification	Method
Escherichia coli ATCC® 25922	Pass	Grey, wet colonies	QC75
Staphylococcus aureus ATCC® 25923	Pass	White colonies	QC75
Streptococcus pyogenes ATCC® 19615	Pass	Beta haemolytic colonies	QC75
Staphylococcus saprophyticus ATCC® 1530	Pass	White colonies	QC75

**Physical Testing Side 2** **Results** **Specification**  
*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Appearance (Colour) - Chrom	Pass	White/Cream
Clarity - Chrom	Pass	Opaque
Gel Strength	Pass	Passed
pH - Chrom	7.0	6.6 - 7.0

Microbiological Testing Side 2	Result	Specification	Method
Escherichia coli ATCC® 25922	Pass	Pink colonies	QC75
Staphylococcus aureus ATCC® 25923	Pass	Cream colonies	QC75
Klebsiella pneumoniae ATCC® 13883	Pass	Blue colonies	QC75
Proteus mirabilis ATCC® 12453	Pass	Beige cols, no swarming	QC75
Enterococcus faecalis ATCC® 29212	Pass	Blue/green colonies	QC75
Staphylococcus saprophyticus ATCC® 1530	Pass	Pink/white colonies	QC75

*Paul Costello*

**Approved Signatory**

Accredited for compliance with ISO/IEC 17025 - Testing.  
This report may not be reproduced except in full.



Accreditation  
Number 3186



Accreditation No. 3186

**MANUFACTURED AND TESTED BY:**  
**Thermo Fisher Scientific, Microbiology Division**  
20 Dalglish Street, Thebarton, South Australia 5031  
Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054  
Email: [micro\\_orders@thermofisher.com](mailto:micro_orders@thermofisher.com) Website: [www.thermofisher.com.au](http://www.thermofisher.com.au)

## CERTIFICATE OF ANALYSIS

---

<b>Microbial Load Testing</b>	<b>Results</b>	<b>Specification</b>
-------------------------------	----------------	----------------------

---

*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

---

Microbial Load @ 30°C for 3 days /sample	Pass	Within Limits
--	------	---------------

*This test is performed in accordance Oxoid method QC21, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1*

---

Tested in accordance with QSP1105

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

**MANUFACTURED AND TESTED BY:**

**Thermo Fisher Scientific, Microbiology Division**

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: [micro\\_orders@thermofisher.com](mailto:micro_orders@thermofisher.com) Website: [www.thermofisher.com.au](http://www.thermofisher.com.au)