

# CERTIFICATE OF ANALYSIS

**PRODUCT** PP2249  
**BRILLIANCE UTI/HBA**

**LOT NUMBER** 4343940

**EXPIRY DATE** 2022.02.02

**RELEASE DATE** 2021.12.17

Physical Testing Side 1	Results	Specification
Appearance (Colour) - HBA	Pass	Red
Clarity - HBA	Pass	Opaque
Gel Strength	Pass	Satisfactory
pH - HBA	Pass	7.2 - 7.6

Microbiological Testing Side 1	Result	Specification
Escherichia coli ATCC® 25922	Pass	Grey, wet colonies
Staphylococcus aureus ATCC® 25923	Pass	White colonies
Streptococcus pyogenes ATCC® 19615	Pass	Beta haemolytic colonies
Staphylococcus saprophyticus ATCC® 1530	Pass	White colonies

Physical Testing Side 2	Results	Specification
Appearance (Colour) - Chrom	Pass	White/Cream
Clarity - Chrom	Pass	Opaque
Gel Strength	Pass	Passed
pH - Chrom	Pass	6.6 - 7.0

Microbiological Testing Side 2	Result	Specification
Escherichia coli ATCC® 25922	Pass	Pink colonies
Staphylococcus aureus ATCC® 25923	Pass	Cream colonies
Proteus mirabilis ATCC® 12453	Pass	Beige cols, no swarming
Enterococcus faecalis ATCC® 29212	Pass	Blue/green colonies
Staphylococcus saprophyticus ATCC® 1530	Pass	Pink/white colonies

Microbial Load Testing	Results	Specification
Microbial Load @ 30°C for 3 days /sample	Pass	AQL 2.5

Tested in accordance with QSP1105 on a representative sample of the batch. Results obtained at the time of release.  
Decision rules are outlined in the product Technical Specification.  
Uncertainty of measurement introduced during quantitative testing has been determined. Values not reported on the certificate of analysis, but details available on request.

Accredited for compliance with ISO/IEC 17025 - Testing.  
This report may not be reproduced except in full.



Accreditation  
Number 3186

**MANUFACTURED AND TESTED BY:**  
**Thermo Fisher Scientific, Oxoid Australia Pty Ltd**  
20 Dalgleish Street, Thebarton, South Australia 5031  
Tel: Toll Free 1300 735 292, Fax: 1800 007 054  
Email: microtechsupport@thermofisher.com  
Website: www.thermofisher.com

# CERTIFICATE OF ANALYSIS

Released by: Anna Catalano



.....  
**Approved Signatory**

Accredited for compliance with ISO/IEC 17025 - Testing.  
This report may not be reproduced except in full.



Accreditation  
Number 3186

**MANUFACTURED AND TESTED BY:**  
**Thermo Fisher Scientific, Oxoid Australia Pty Ltd**  
20 Dalgleish Street, Thebarton, South Australia 5031  
Tel: Toll Free 1300 735 292, Fax: 1800 007 054  
Email: [microtechsupport@thermofisher.com](mailto:microtechsupport@thermofisher.com)  
Website: [www.thermofisher.com](http://www.thermofisher.com)