

CERTIFICATE OF ANALYSIS

PRODUCT	PP2269 CHROMOGENIC SALMONELLA
LOT NUMBER	4277551
EXPIRY DATE	2020.10.29

RELEASE DATE 2020.09.06

Physical TestingResultsSpecificationTested in accordance with Oxoid Method QC13 unless otherwise stated							
Appearance (Colour)	Pass	White/Pale Pink					
Clarity	Pass	Opaque					
Gel Strength	Pass	Satisfactory					
pH	7.4	7.0 - 7.4					
Microbial Load @ 30°C for 3 days /sample	Pass	Within Limits					
This test is performed in accordance Oxoid metho General inspection level=1	od QC21, ASM (Guidelines: Normal inspection AQL 2.5,					

Microbiology Testing	Result	Specification	Method	
Salmonella typhimurium ATCC® 14028	Pass	Pink centres, clear edges	QC75	
Escherichia coli ATCC® 25922	Pass	Pale blue colonies	QC75	
Pseudomonas aeruginosa ATCC® 27853	Pass	Inhibited	QC75	
Salmonella hofit IMVS 1799	Pass	Pink colonies	QC75	

Tested in accordance with QSP1105

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Amy-Lee Maggs

.....

Approved Signatory

Accredited for compliance with ISO/IEC 17025 - Testing. This report may not be reproduced except in full.





MANUFACTURED AND TESTED BY: Thermo Fisher Scientific, Microbiology Division 20 Dalgleish Street, Thebarton, South Australia 5031 Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054 Email: micro_orders@thermofisher.com Website: www.thermofisher.com.au

Number 3186