

CERTIFICATE OF ANALYSIS

PRODUCT PP2269
CHROMOGENIC SALMONELLA

LOT NUMBER 4288375

EXPIRY DATE 2021.01.13

RELEASE DATE 2020.11.22

Physical Testing	Results	Specification
------------------	---------	---------------

Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	White/Pale Pink
Clarity	Pass	Opaque
Gel Strength	Pass	Satisfactory
pH	7.4	7.0 - 7.4
Microbial Load @ 30°C for 3 days /sample	Pass	Within Limits

This test is performed in accordance Oxoid method QC21, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1

Microbiology Testing	Result	Specification	Method
----------------------	--------	---------------	--------

Salmonella typhimurium ATCC® 14028	Pass	Pink centres, clear edges	QC75
Escherichia coli ATCC® 25922	Pass	Pale blue colonies	QC75
Pseudomonas aeruginosa ATCC® 27853	Pass	Inhibited	QC75
Salmonella hofit IMVS 1799	Pass	Pink colonies	QC75

Tested in accordance with QSP1105

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Paul Costello

.....
Approved Signatory

Accredited for compliance with ISO/IEC 17025 - Testing.
This report may not be reproduced except in full.



Accreditation
Number 3186



Accreditation No. 3186

MANUFACTURED AND TESTED BY:

Thermo Fisher Scientific, Microbiology Division

20 Dalgleish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: micro_orders@thermofisher.com Website: www.thermofisher.com.au