

**CERTIFICATE OF ANALYSIS**

**PRODUCT** PP2269  
**CHROMOGENIC SALMONELLA**

**LOT NUMBER** 4340072

**EXPIRY DATE** 2022.01.13

**RELEASE DATE** 2021.11.21

Physical Testing	Results	Specification
Appearance (Colour)	Pass	White/Pale Pink
Clarity	Pass	Opaque
Gel Strength	Pass	Satisfactory
pH	Pass	7.0 - 7.4
Microbial Load @ 30°C for 3 days /sample	Pass	AQL 2.5


Microbiology Testing	Result	Specification
Salmonella typhimurium ATCC® 14028	Pass	Pink centres, clear edges
Escherichia coli ATCC® 25922	Pass	Pale blue colonies
Pseudomonas aeruginosa ATCC® 27853	Pass	Inhibited
Salmonella hofit IMVS 1799	Pass	Pink colonies

Tested in accordance with QSP1105 on a representative sample of the batch. Results obtained at the time of release.

Decision rules are outlined in the product Technical Specification.

Uncertainty of measurement introduced during quantitative testing has been determined. Values not reported on the certificate of analysis, but details available on request.

Released by: Paul Costello



.....  
**Approved Signatory**

Accredited for compliance with ISO/IEC 17025 - Testing.  
This report may not be reproduced except in full.



Accreditation  
Number 3186

**MANUFACTURED AND TESTED BY:**  
**Thermo Fisher Scientific, Oxoid Australia Pty Ltd**  
20 Dalgleish Street, Thebarton, South Australia 5031  
Tel: Toll Free 1300 735 292, Fax: 1800 007 054  
Email: [microtechsupport@thermofisher.com](mailto:microtechsupport@thermofisher.com)  
Website: [www.thermofisher.com](http://www.thermofisher.com)