

**CERTIFICATE OF ANALYSIS**

**PRODUCT** PP2272  
**CHROMOGENIC CANDIDA**

**LOT NUMBER** 4312170

**EXPIRY DATE** 2021.07.08

**RELEASE DATE** 2021.05.05

**Physical Testing Results Specification**

*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Physical Testing	Results	Specification
Appearance (Colour)	Pass	Colourless
Clarity	Pass	Translucent
Gel Strength	Pass	Satisfactory
pH	6.4	6.0 - 6.4
Microbial Load@ 23°C for 7days/ sample	Pass	Within Limits

*This test is performed in accordance Oxoid method QC21, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1*

**Microbiology Testing Result Specification Method**

Microbiology Testing	Result	Specification	Method
Candida albicans ATCC® 10231	Pass	Green colonies	QC75
Candida tropicalis ATCC® 0750	Pass	Blue colonies	QC75
Candida krusei ATCC® 14243	Pass	Irregular pink colonies	QC75
Candida glabrata ATCC® 2001	Pass	Dark pink colonies	QC75
Candida parapsilosis ATCC® 22019	Pass	Pale pink/Cream colonies	QC75

Tested in accordance with QSP1105

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

*Paul Costello*

**Approved Signatory**

Accredited for compliance with ISO/IEC 17025 - Testing.  
This report may not be reproduced except in full.



Accreditation  
Number 3186



Accreditation No. 3186

**MANUFACTURED AND TESTED BY:**

**Thermo Fisher Scientific, Microbiology Division**

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: [micro\\_orders@thermofisher.com](mailto:micro_orders@thermofisher.com) Website: [www.thermofisher.com.au](http://www.thermofisher.com.au)