

CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PP2272

CHROMOGENIC CANDIDA

LOT NUMBER 4336858

EXPIRY DATE 2022.01.05

RELEASE DATE 2021.11.02

Physical Testing	Results	Specification
Appearance (Colour)	Pass	Colourless
Clarity	Pass	Translucent
Gel Strength	Pass	Satisfactory
pH	Pass	6.0 - 6.4
Microbial Load@ 23°C for 7days/ sample	Pass	AQL 2.5
Microbiology Testing	Result	Specification
Candida albicans ATCC® 10231	Pass	Green colonies
Candida tropicalis ATCC® 0750	Pass	Blue colonies
Candida krusei ATCC® 14243	Pass	Irregular pink colonies
Candida glabrata ATCC® 2001	Pass	Dark pink colonies
Candida parapsilosis ATCC® 22019	Pass	Pale pink/Cream colonies

Tested in accordance with QSP1105 on a representative sample of the batch. Results obtained at the time of release.

Decision rules are outlined in the product Technical Specification.

Uncertainty of measurement introduced during quantitative testing has been determined. Values not reported on the certificate of analysis, but details available on request.

Released by: Paul Costello

Paul Costolo

Approved Signatory

Accredited for compliance with ISO/IEC 17025 - Testing. This report may not be reproduced except in full.



MANUFACTURED AND TESTED BY: Thermo Fisher Scientific, Oxoid Australia Pty Ltd 20 Dalgleish Street, Thebarton, South Australia 5031 Tel: Toll Free 1300 735 292, Fax: 1800 007 054 Email: microtechsupport@thermofisher.com

Website: www.thermofisher.com