

Certificate of Quality

Product Name: Anaer Egg Yolk Agar 10/PK
Lot Number: 872473

Product Number: R01057
Expiration Date: 2019-12-11
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Clostridium perfringens	13124	Good Growth, Lipase (-), Lecithinase (+)
Clostridium sporogenes	3584	Good Growth, Lipase (+), Lecithinase (-)

Physical Characteristics:

Appearance: Light Amber, Opaque Agar
pH @ 20-25°C: 7.0 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor