

Certificate of Quality

Product Name: Anaer Reducible Blood Agar 15/PK
Lot Number: 324254

Product Number: R01061
Expiration Date: 2018-07-04
 (YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|-------------------------------|-------------|-----------------------------|
| Bacteroides fragilis | 25285 | Good Growth |
| Clostridium perfringens | 13124 | Good Growth, Beta Hemolysis |
| Fusobacterium nucleatum | 25586 | Good Growth |
| Peptostreptococcus anaerobius | 27337 | Good Growth |
| Prevotella melaninogenica | 25845 | Good Growth |
| Escherichia coli | 25922 | Good Growth |
| Staphylococcus aureus | 25923 | Good Growth |

Physical Characteristics:

Appearance: Red, Opaque Agar
 pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor