

## Certificate of Quality

**Product Name:** Anaer Reducible CNA Bld Agar 15/PK  
**Lot Number:** 459463

**Product Number:** R01065  
**Expiration Date:** 2019-03-14  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

| <b>Organism Challenge</b>     | <b>ATCC</b> | <b>Performance Criteria</b> |
|-------------------------------|-------------|-----------------------------|
| Bacteroides fragilis          | 25285       | Good Growth                 |
| Peptostreptococcus anaerobius | 27337       | Good Growth                 |
| Staphylococcus aureus         | 25923       | Good Growth                 |
| Escherichia coli              | 25922       | Inhibited                   |

**Physical Characteristics:**

Appearance: Red, Opaque Agar  
pH @ 20-25°C: 7.3 +/- 0.2

**Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor