

Certificate of Quality

Product Name: BHI Agar w/SB, Pen, Strep, Cyclo(25ml)10/PK
Lot Number: 275675


Product Number: R01156
Expiration Date: 2021-06-29
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|-----------------------------|-------|----------------------|
| Candida albicans | 10231 | Good Growth |
| Trichophyton mentagrophytes | 9533 | Good Growth |
| Aspergillus niger | 16404 | Inhibited |
| Cryptococcus neoformans | 14116 | No Growth |
| Escherichia coli | 25922 | No Growth |
| Staphylococcus aureus | 25923 | No Growth |

Physical Characteristics:
Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.4 +/- 0.2

Microbial Load Evaluation:
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Senior Quality Assurance Engineer