

Certificate of Quality

Product Name: BHI Agar w/SB, Pen, Strep, Cyclo(25ml)10/PK
Lot Number: 967029


Product Number: R01156
Expiration Date: 2020-07-08
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Candida albicans	10231	Good Growth
Trichophyton mentagrophytes	9533	Good Growth
Aspergillus niger	16404	Inhibited
Cryptococcus neoformans	14116	No Growth
Escherichia coli	25922	No Growth
Staphylococcus aureus	25923	No Growth

Physical Characteristics:
Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.4 +/- 0.2

Microbial Load Evaluation:
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor