

Certificate of Quality

Product Name: BHI Agar w/Vanc (6ug) 10/PK
Lot Number: 435866

Product Number: R01176
Expiration Date: 2019-01-10
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Enterococcus faecalis	29212	No Growth
Enterococcus faecalis	49532	No Growth
Enterococcus faecalis	49533	No Growth
Enterococcus faecalis	51299	Good Growth
Staphylococcus aureus	25923	No Growth

Physical Characteristics:

Appearance: Light To Medium Amber, Clear Agar
pH @ 20-25°C: 7.4 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor