

Certificate of Quality

Product Name: Brucella Agar w/Horse Blood,TVP 10/PK
Lot Number: 215437

Product Number: R01261
Expiration Date: 2021-02-23
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Campylobacter jejuni	33291	Good Growth
Escherichia coli	25922	No Growth
Proteus mirabilis	12453	Inhibited, Swarming
Staphylococcus aureus	25923	No Growth

Physical Characteristics:

Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.0 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor