

Certificate of Quality

Product Name: Chocolate Agar 100/PK
Lot Number: 431621

Product Number: R01302
Expiration Date: 2019-01-21
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Haemophilus influenzae	10211	Good Growth
Neisseria gonorrhoeae	43069	Good Growth
Neisseria meningitidis	13090	Good Growth
Streptococcus pneumoniae	6305	Good Growth

Physical Characteristics:

Appearance: Brown, Opaque Agar
pH @ 20-25°C: 7.2 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor