

Certificate of Quality

Product Name: HE Agar (Hektoen Enteric) Agar 15/PK
Lot Number: 196705

Product Number: R01481
Expiration Date: 2021-02-15
 (YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Salmonella typhimurium	14028	Green Colonies, Black Centers
Shigella flexneri	12022	Green Colonies
Enterococcus faecalis	29212	No Growth
Escherichia coli	25922	Inhibited, Yellow Colonies

Physical Characteristics:

Appearance: Green, Clear Agar
 pH @ 20-25°C: 7.6 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor