

## Certificate of Quality

**Product Name:** MacConkey w/Sorbitol 10/PK  
**Lot Number:** 269713

**Product Number:** R01556  
**Expiration Date:** 2013-01-21  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Escherichia coli	25922	Pink Colonies
Escherichia coli	43888	Clear Colonies
Proteus mirabilis	12453	Clear Colonies, No Swarming
Salmonella enterica	14028	Pink Colonies
Staphylococcus aureus	25923	No Growth
Enterococcus faecalis	29212	No Growth

**Physical Characteristics:**

Appearance: Pink, Clear Agar  
pH @ 20-25°C: 7.1 +/- 0.2

**Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor