

Certificate of Quality

Product Name: V Agar, Selective 10/PK
Lot Number: 333431

Product Number: R01952
Expiration Date: 2018-06-20
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|---------------------------|-------------|-----------------------------|
| Gardnerella vaginalis | 14018 | Good Growth, Beta Hemolysis |
| Escherichia coli | 25922 | No Growth |
| Proteus mirabilis | 12453 | No Growth |
| Candida albicans | 10231 | No Growth |

Physical Characteristics:

Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.2 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor