

Certificate of Quality

Product Name: Blood/Chocolate Biplate 100/PK
Lot Number: 462456

Product Number: R02029
Expiration Date: 2019-03-18
 (YYYY-MM-DD)


This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Blood Agar:		
Streptococcus pyogenes	19615	Good Growth, Beta Hemolysis
Streptococcus agalactiae	12386	Good Growth, Beta Hemolysis
Enterococcus faecalis	29212	Good Growth, Gamma Hemolysis
Streptococcus pneumoniae	6305	Good Growth, Alpha Hemolysis
Listeria monocytogenes	7646	Good Growth, Beta Hemolysis
Escherichia coli	25922	Good Growth
Staphylococcus aureus	25923	Good Growth
Moraxella catarrhalis	25238	Good Growth

Chocolate Agar:		
Haemophilus influenzae	10211	Good Growth
Neisseria meningitidis	13090	Good Growth
Neisseria gonorrhoeae	43069	Good Growth
Streptococcus pneumoniae	6305	Good Growth

Physical Characteristics:
 Appearance: Blood Agar: Red, Opaque Agar
 Chocolate agar: Brown, Opaque Agar
 pH @ 20-25°C: N/A

Microbial Load Evaluation:
 Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

 Quality Assurance Supervisor