

Certificate of Quality

Product Name: GC Agar Base (150mm) 10/PK
Lot Number: 358417

Product Number: R04030
Expiration Date: 2021-11-17
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

This product has been tested and is compliant with the performance criteria in the NCCLS document, "Performance Standards For Antimicrobial Disk Susceptibility Tests."


Physical Characteristics:

Appearance: Light to Medium Amber, Clear Agar
pH @ 20-25°C: 7.2 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Senior Quality Assurance Engineer