

Certificate of Quality

Product Name: HI Broth w/BCP,1% Maltose 20/PK
Lot Number: 533545

Product Number: R061100
Expiration Date: 2020-05-15
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Streptococcus bovis	9809	(+) Yellow
Morganella morganii	25830	(-) No Color Change

Physical Characteristics:

Appearance: Dark Purple, Clear Broth
pH @ 20-25°C: 7.4 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor