

Certificate of Quality

Product Name: Chocolate Agar (slant) 20/PK
Lot Number: 172192

Product Number: R08240
Expiration Date: 2018-06-19
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
<i>Haemophilus influenzae</i>	10211	Good Growth
<i>Neisseria gonorrhoeae</i>	43069	Good Growth
<i>Neisseria meningitidis</i>	13090	Good Growth
<i>Streptococcus pneumoniae</i>	6305	Good Growth

Physical Characteristics:

Appearance: Brown, Opaque Agar Slant
pH @ 20-25°C: 7.2 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor