

## Certificate of Quality

**Product Name:** LJ Medium (1ozBTL) 10/PK  
**Lot Number:** 334968

**Product Number:** R10101  
**Expiration Date:** 2022-07-22  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Mycobacterium fortuitum	6841	Good Growth
Mycobacterium kansasii	12478	Good Growth
Mycobacterium intracellulare	13950	Good Growth
Mycobacterium scrofulaceum	19981	Good Growth
Mycobacterium tuberculosis	25177	Good Growth


**Physical Characteristics:**

Appearance: Pale Green, Opaque Agar Slant  
pH @ 20-25°C: N/A

**Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Senior Quality Assurance Engineer