

Certificate of Quality

Product Name: LJ Medium (1ozBTL) 10/PK
Lot Number: 853728

Product Number: R10101
Expiration Date: 2020-09-03
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Mycobacterium fortuitum	6841	Good Growth
Mycobacterium kansasii	12478	Good Growth
Mycobacterium intracellulare	13950	Good Growth
Mycobacterium scrofulaceum	19981	Good Growth
Mycobacterium tuberculosis	25177	Good Growth

Physical Characteristics:

Appearance: Pale Green, Opaque Agar Slant
pH @ 20-25°C: N/A

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor