

Certificate of Quality

Product Name: Brill Green Agar w/Novobiocin 100/PK
Lot Number: 157988

Product Number: R110113
Expiration Date: 2020-12-16
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Salmonella enterica	14028	Good Growth with Pink Colonies & Zone
Proteus mirabilis	12453	No Growth

Physical Characteristics:

Appearance: Green, Clear Agar
pH @ 20-25°C: 6.9 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor