

Certificate of Quality

Product Name: XLT4/Bril Green w/Novo Biplate 100/PK
Lot Number: 420219

Product Number: R110506
Expiration Date: 2018-12-24
 (YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
BG with Novo:		
Salmonella typhimurium	14028	Pink Colonies w/ Pink Zone
Proteus mirabilis	12453	No Growth
XLT4:		
Salmonella typhimurium	14028	Red Colonies w/ Black Centers
Proteus mirabilis	12453	No Growth
Citrobacter freundii	33128	Yellow Colonies

Physical Characteristics:

Appearance:	Brilliant Green Side: Green, Clear Agar XLT4 Side: Red-Orange, Clear Agar
pH @ 20-25°C:	N/A

Microbial Load Evaluation:
 Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

 Quality Assurance Supervisor