

Certificate of Quality

Product Name: Contact I Sab Dex w/Lec,PS80 20/PK
Lot Number: 158144

Product Number: R111105
Expiration Date: 2012-05-09
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Results
Candida albicans *	10231	Pass
Cryptococcus neoformans*	34877	Pass
Trichophyton mentagrophytes	9533	Pass

A pass result indicates organism produced the expected colony size, number, and morphology, indicative of the challenge organism and level for the media formulation, within the required time period. Inoculation levels of 10-100 CFUs were used for the indicated (*) organisms, and incubation was conducted at 20-25°C for not more than three days. The results meet or exceed USP Growth Promotion requirements.

Physical Characteristics:

Appearance: Light to Medium Amber, Clear Agar
pH @ 20-25°C: 5.6 +/- 0.2

Microbial Load Evaluation:

Product has been manufactured in a particulate and microbiologically controlled environment. Product was statistically sampled to an AQL of 0.25. Samples were acceptable for microbial content and aesthetic value following 7 days at 20-25°C.

Certificate results were obtained at time of release. End users are encouraged to assure product meets internal standards at time of use.

Signed



Quality Assurance Supervisor