

Certificate of Quality

Product Name: M4RT (3ml) w/o BEADS PK/72
Lot Number: 196365

Product Number: R126990
Expiration Date: 2022-06-02
 (YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	Strain	Results
---------------------------	---------------	----------------

Maintenance:

Chlamydia trachomatis	VR-880	Pass
Herpes simplex type 2	VR-734	Pass
Cytomegalovirus	VR-977	Pass
Respiratory syncytial virus	VR-1540	Pass

Toxicity:

MRC5 Cells	CCL-171	Pass
VERO Cells	CCL-81	Pass
McCoy Cells	CRL-1696	Pass

Physical Characteristics:

Appearance:	Orange-Pink to Pink-Orange Clear Broth
pH @ 20-25°C:	7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following half of the samples incubated at a minimum of 72 hours at 33-37°C, subcultured to Blood Agar, and incubated at an additional 72 hours at 33-37°C, and the other half of the samples incubated at a minimum of 48 hours at 20-25°C, subcultured to Sabouraud Dextrose Agar, and incubated at an additional 5 days at 20-25°C.

Signed



Quality Assurance Supervisor