

Certificate of Quality

Product Name: A7 Agar (60mm) 10/PK
Lot Number: 251120

Product Number: R20201
Expiration Date: 2012-11-15
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Mycoplasma hominis	23114	Good Growth
Ureaplasma urealyticum	27618	Good Growth, Golden Brown Colonies
Staphylococcus aureus	25923	No Growth
Candida albicans	10231	No Growth

Physical Characteristics:

Appearance: Light To Medium Amber, Clear Agar
pH @ 20-25°C: 6.0 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor