

Certificate of Quality

Product Name: 10B Arginine Broth (1.8ml) 100/PK
Lot Number: 313951

Product Number: R20305
Expiration Date: 2019-03-26
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Ureaplasma urealyticum	27618	Pink Color Change, Good Recovery, Golden Brown Colonies
Mycobacterium hominis	23114	Red Color Change, Good Recovery

Physical Characteristics:

Appearance: Yellow, Clear Broth
pH @ 20-25°C: 6.0 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor