

Certificate of Analysis

Product Name: ProSpecT Shiga Toxin (STEC) MP KT/96TST
Product Code: R2474096
Lot Number: 2908229
Expiration Date: 28.02.2021
Date of Manufacture: 13.03.2020

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74PCJ05A	Shiga Toxin Conjugate 25ml Filled Bulk	2894494	28.02.2021
W74PCS01P	Colour Substrate TMB-W 25ml Filled Bulk	2894539	31.07.2021
W74PDL01T	Diluent 110ml SLT Filled Bulk	2894497	28.02.2022
W74PNC03T	Bacterial Negative Control Filled Bulk	2894496	28.02.2022
W74PPC05H	Shiga Toxin Positive Control Filled Bulk	2894495	28.02.2022
W74PPLO5E	SLT I&II Plate 96 Test	2500406	30.09.2021
W74PSP01A	Stop Solution Filled Bulk	2894540	31.05.2021
W74PWS01C	Wash Solution 110ml Filled Bulk	2894498	28.02.2022

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Specification	Result
Negative Control	Visually Negative	Pass
	Mean OD < 0.100 (s)	Pass
	Mean OD < 0.070 (d)	Pass
Positive Control	Visually Positive	Pass
	Mean OD > 0.500 (s,d)	Pass
Positive Samples	Visually Positive	Pass
	Mean OD > 0.150 (s)	Pass
	Mean OD > 0.100 (d)	Pass
Negative Samples	Visually Negative	Pass
	Mean OD < 0.100 (s)	Pass
	Mean OD < 0.070 (d)	Pass



Fay Curd
Quality Assurance Manager