

## Certificate of Analysis

**Product Name:** Thrombo-Wellcotest KT/20TST  
**Product Code:** R30852601  
**Lot Number:** 2263667  
**Expiration Date:** 2019-02-28  
**Date of Manufacture:** 2018-01-09

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74CHA13T	HA13 Glycine Buffer Filled Bulk	2175399	2019-08-31
W74MHA13Q	Thrombo Neg Control Filled Bulk	2175398	2019-02-28
W74YHA13A	Thrombo Pos Control	2175397	2019-02-28
W75EHA08F	Thrombo Latex filled bulk	2229603	2019-02-28
W7CAHA25D	Haemoguard Vacutainer	7150910	2019-05-31

Test	Required result	Actual Result
------	-----------------	---------------

Latex sensitivity with a Standard FDP serum	The 2 micrograms/ml FDP Standard Serum must show at least + agglutination with the Latex	Conforms
---	--	----------

	The 1 microgram/ml FDP Standard Serum must not show agglutination although a trace is acceptable.	Conforms
--	---	----------

Positive & Negative Control sera check	The Positive Control Serum must show at least ++ agglutination with the Latex.	Conforms
--	--	----------

	The Negative Control Serum must not show any agglutination with the Latex although a trace is acceptable.	Conforms
--	---	----------



Fay Curd  
Quality Assurance Manager