

Certificate of Analysis

Product Name: Salm paratyphi A-O stain susp VL/5ml
Product Code: R30953301
Lot Number: 1904606
Expiration Date: 2018-10-31
Date of Manufacture: 2016-11-08

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74KSS11K	S.paratyphi A	1904607	2018-10-31

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Anticipated result	Actual Result
Efficacy in rapid slide titration	The titre obtained with the antiserum must be roughly equivalent to the titre obtained in the tube agglutination test	Conforms
	There must be no agglutination in the saline control	Conforms
Efficacy in tube agglutination	The Stained Suspension under test must show agglutination to the titre stated on the antiserum bottle plus or minus two doubling dilutions	Conforms
	There must be no agglutination in the saline control	Conforms
	A Reference batch of Stained Suspension must agglutinate to the titre stated on the antiserum bottle plus or minus two doubling dilutions	Conforms



Fay Curd
Quality Assurance Manager