

Certificate of Quality

Product Name: GC Agar Base 10kg each
Lot Number: 712768

Product Number: R453506
Expiration Date: 2019-08-01
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Haemophilus influenzae	10211	Good Growth
Neisseria meningitidis	13090	Good Growth
Neisseria gonorrhoeae	43069	Good Growth
Streptococcus pneumoniae	6305	Good Growth

Visual Driform: Beige, homogeneous, free-flowing powder

pH: Upon rehydration per manufacturer specifications, media has a pH of 7.2 +/- 0.2

This lot meets criteria for approval as defined for this product. It should be stable to the expiration date on the label when stored according to package specifications.

Signed



Quality Assurance Supervisor