

Certificate of Quality

Product Name: M Broth 2.5kg each
Lot Number: 353453

Product Number: R453784
Expiration Date: 2017-05-25
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Salmonella enterica	14028	Good Growth
Salmonella enterica	13312	Good Growth
Salmonella enterica	19430	Good Growth
Salmonella enterica	6539	Good Growth
Salmonella enterica	13314	Good Growth

Visual Driform: Light beige, homogeneous, free-flowing powder

pH: Upon rehydration per manufacturer specifications, media has a pH of 7.0 +/- 0.2

This lot meets criteria for approval as defined for this product. It should be stable to the expiration date on the label when stored according to package specifications.

Signed



Quality Assurance Supervisor