thermoscientific

Thermo Fisher Scientific Microbiology 12076 Santa Fe Trail Drive 12230 Santa Fe Trail Drive Lenexa, KS 66215 800.255.6730 800.447.5761 fax www.thermofisher.com

Certificate of Analysis

Product Name: P. aeruginosa ATCC 9027 Q KT/10 TST Lot Number: 102326

Product Number: R4735210 Expiration Date: 2021-07-08 (YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Purity:

Standardized aliquots of the rehydrated product are inoculated onto nonselective media and examined for pure growth following the appropriate incubation. Selective and Differential media are also tested where applicable.

Viability And Quantification:

Each organism is recovered from the preserved state within the required time frame and at an acceptable level. Passage number is stated as the current preserved state.

Macroscopic And Microscopic Morphology:

Colony morphology is consistent with documented referenced description. Traditional staining is performed.

Characterization:

Organism exhibits characteristic biochemical, enzymatic, genotypical and/or biochemical reactions. Automated and/or conventional testing was performed and results were within established limits. Antimicrobial testing performed where applicable. Results within expected ranges.

CFU/ 0.3ml:<100</th>Passage:2Gram Reaction:Gram Negative RodBiochemical Profile:Vitek 2C GN

Appearance: Preserved Gel Matrix suspended in vial cap pH: N/A

Colony Count: Not less than 5 samples each were plated to the following: Trypticase Soy Agar, spread plates, incubated at 35°C for 24 hours.

Colony forming units as noted above are specified under defined growing conditions. Results may vary with more inhibitory or selective media or with the same medium of inferior quality.

CFU Mean 51



Certificate of Analysis

Eigne Willians

Quality Assurance Manager

The identity, purity, and authenticity of the Licensed Products are exclusively the responsibility of Remel Inc. and not ATCC. The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative Word mark, and the ATCC Catalog Marks are trademarks of ATCC.

Remel Inc. is licensed to use these trademarks and to sell products derived from ATCC® culture.

