

Certificate of Analysis

Product Name: PathoDx Herpes Typing KT/100TST
Product Code: R62250
Lot Number: 2299362
Expiration Date: 2018-12-31
Date of Manufacture: 2018-03-28

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74IFA01A	Mounting Fluid 5ml Filled Bulk	2286815	2019-07-31
W74IFA18C	HSV 1 Rgt 4.5ml	2248229	2019-06-30
W74IFA19D	HSV II Rgt 4.5ml	2248230	2019-06-30
W74IFA23J	Herpes Slides (pouched and labelled)	2248234	2018-12-31

Test results carried out according to the Quality Assurance Product Release and Test Specification:

Test	Required Result	Actual Result
Kit Compatibility Test	The HSV-I Typing Reagent must show fluorescence with the homologous HSV-1 well on the Herpes Typing Control Slide and show no fluorescence with the heterologous HSV-II well	Conforms
	The HSV-II Typing Reagent must show fluorescence with the homologous HSV-II well on the Herpes Typing Control Slide and show no fluorescence with the heterologous HSV-I well	Conforms



Fay Curd
Quality Assurance Manager