



Quality Assurance Certificate

Code: SP0409S
Product: SENSITEST AGAR 10 x 9L (10 x 288g)
Lot: 1886310
Expiry: 31 July 2021
Components:

PHYSICAL CHARACTERISTICS: As appropriate tests on moisture, appearance, pH, clarity and gel strength were satisfactory.

BIOLOGICAL PERFORMANCE: Satisfactory performance of:

Streptococcus pyogenes	ATCC 19615
Streptococcus pneumoniae	ATCC 6303
Haemophilus influenzae	ATCC 19418
Staphylococcus aureus	NCTC 6571
Enterococcus faecalis	ATCC 29212
Staphylococcus aureus	ATCC 25923
Escherichia coli	ATCC 25922
Pseudomonas aeruginosa	ATCC 27853
Proteus mirabilis	NCTC 10975
Enterococcus faecalis	ATCC 33186

Date: 04 July 2018

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct, all results reported in this certificate relate only to the product and lot stated in this certificate of analysis. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual.

Shannon Bearpark

Shannon Bearpark
Site Quality Manager

