

**CERTIFICATE OF ANALYSIS**

**PRODUCT** TM4131  
TSB 10ML

**LOT NUMBER** 1978787

**EXPIRY DATE** 2018.10.10

**RELEASE DATE** 2017.10.24

**Physical Testing Results Specification**

*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Appearance (Colour)	Pass	Beige/brown
Clarity	Pass	Translucent
pH	7.4	7.1 - 7.5
Sterility @ 35°C for 14days /sample plan	Pass	Within Limits
<i>This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1</i>		
Sterility @ RT for 14 days /sample plan	Pass	Within Limits
<i>This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5, General inspection level=1</i>		

<b>Microbiological Testing</b>	<b>Result</b>	<b>Specification</b>	<b>Method</b>
Candida albicans ATCC® 10231	Pass	Growth	QC27
Bacillus subtilis ATCC® 6633	Pass	Growth	QC27
Aspergillus brasiliensis ATCC® 16404	Pass	Growth	QC27

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

The product is stable to the stated expiry date when stored in accordance with instructions per product label.

Released by: Paul Costello

*Paul Costello*

**Approved Signatory**

Accredited for compliance with ISO/IEC 17025.  
This report may not be reproduced except in full.



**MANUFACTURED AND TESTED BY:**  
**Thermo Fisher Scientific, Microbiology Division**  
 20 Dalgleish Street, Thebarton, South Australia 5031  
 Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054  
 Email: micro\_orders@thermofisher.com Website: www.thermofisher.com.au