

# CERTIFICATE OF ANALYSIS

**PRODUCT** TM4205  
DERMASEL + ANTI + ACTI 12ML (TRAY 75)

**LOT NUMBER** 2177642

**EXPIRY DATE** 2018.09.04

**RELEASE DATE** 2018.03.08

---

**Physical Testing** **Results** **Specification**  
*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

---

Appearance (Colour)	Pass	Beige/Pale Brown
Gel Strength	Pass	Satisfactory
Clarity	Pass	Translucent
pH	6.8	6.4 - 6.8

---

<b>Microbiological Testing</b>	<b>Result</b>	<b>Specification</b>	<b>Method</b>
Escherichia coli ATCC® 25922	Pass	Inhibited	QC27
Candida albicans ATCC® 10231	Pass	Growth, cream cols	QC03
Aspergillus brasiliensis ATCC® 16404	Pass	Inhibited	QC03
Trichophyton mentagrophytes MVQC 0104	Pass	Growth, white fluffy / downy	QC03
Saccharomyces cerevisiae ATCC® 2601	Pass	Inhibited	QC03

---

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Paul Costello

*Paul Costello*

---

**Approved Signatory**

Accredited for compliance with ISO/IEC 17025.  
This report may not be reproduced except in full.



Accreditation  
Number 3186



Accreditation No. 3186

**MANUFACTURED AND TESTED BY:**

**Thermo Fisher Scientific, Microbiology Division**

20 Dalglish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax Toll free 1 800 007 054

Email: [micro\\_orders@thermofisher.com](mailto:micro_orders@thermofisher.com) Website: [www.thermofisher.com.au](http://www.thermofisher.com.au)