

CERTIFICATE OF ANALYSIS

PRODUCT TM4601
LAURYL TRYPTOSE S/S 10ML (TRAY 75)

LOT NUMBER 2524398

EXPIRY DATE 2019.05.21

RELEASE DATE 2018.11.20

Physical Testing **Results** **Specification**
Tested in accordance with Oxoid Method QC13 unless otherwise stated

Appearance (Colour)	Pass	Yellow / Durham present
Clarity	Pass	Translucent
pH	6.8	6.7 - 7.1

Microbiological Testing	Result	Specification	Method
Escherichia coli ATCC® 11775	Pass	Turbid with gas production	QC27
Enterococcus faecalis ATCC® 29212	Pass	No Growth	QC27
Citrobacter freundii ATCC® 8090	Pass	Turbid with gas production	QC27

Tested in accordance with QSP1105

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

The product is stable to the stated expiry date when stored in accordance with instructions per product label.

Paul Costello

Approved Signatory

Accredited for compliance with ISO/IEC 17025 - Testing.
This report may not be reproduced except in full.



MANUFACTURED AND TESTED BY:
Thermo Fisher Scientific, Microbiology Division
 20 Dalglish Street, Thebarton, South Australia 5031
 Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054
 Email: micro_orders@thermofisher.com Website: www.thermofisher.com.au