

CERTIFICATE OF ANALYSIS

PRODUCT TV5016D MAXIMUM RECOVERY DILUENT 9ML

LOT NUMBER 2248140

EXPIRY DATE 2018.11.15

General Characteristics	Results		Specification
Colour	Conforms		Colourless
Appearance	Conforms		Transparent
pН	7.1		6.8 -7.2
Packaging / Presentation	Conforms		Label & Print check
Filling volume / weight (g)	Conforms		8.82 - 9.18
Sterility @ 25 & $36 \pm 1^{\circ}$ C for 72 hours	Conforms		Within Limits
Microbiological Performance	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC®25923 WDCM 00034	96	92	cfu ± 30% of original count
Escherichia coli ATCC®25922 WDCM 00013	84	85	cfu \pm 30% of original count

Biological performance tested in accordance with current version of ISO 11133.

The formulation of this medium conforms to ISO 6887-1 and DIN 10162.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

