

CERTIFICATE OF ANALYSIS

PRODUCT TV5016D MAXIMUM RECOVERY DILUENT 9ML
LOT NUMBER 2270203
EXPIRY DATE 2018.12.27

General Characteristics	Results	Specification
Colour	Conforms	Colourless
Appearance	Conforms	Transparent
pH	7.0	6.8 -7.2
Packaging / Presentation	Conforms	Label & Print check
Filling volume / weight (g)	Conforms	8.82 - 9.18
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC®25923 WDCM 00034	82	82	cfu ± 30% of original count
Escherichia coli ATCC®25922 WDCM 00013	109	97	cfu ± 30% of original count

Biological performance tested in accordance with current version of ISO 11133.

The formulation of this medium conforms to ISO 6887-1 and DIN 10162.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133:2014.