Fisher Diagnostics
a division of Fisher Scientific Company, LLC
a part of Thermo Fisher Scientific Inc.
8365 Valley Pike
Middletown, VA (USA) 22645-1905
Tel: 540 869 3200 or 800-528-0494

Certificate of Analysis

Product Name: GTTB GlucoCrush Orange 50, 24 PK
Product Number: B2495-7A
Lot Number: 162000

Expiration Date: 03/31/2014
Storage: 15 - 30°C

Specification	Specification Limit	Result
pH	< 4.00	3.14
Fill Volume	9.5 - 10.5 oz/bottle	9.9
Dextrose Concentration	46.3 - 53.8 g/bottle	50.1
Bioburden	<= 1 CFU/mL	Complies
Coliform	0 Coliform/100 mL	Complies

For in-vitro diagnostic use.

Dextrose Identification
Specification: Dextrose
Result: Dextrose

Quality Representative Date: 2012-03-10

PRINTED ON: 2012-03-10