

***Certificate of Analysis***

Product Name: ARC Ant-TPO Reagent Kit 1x100  
OEM Part Number: 02K47-25  
OEM Lot Number: 26732UN16  
Storage: 2 - 8°C  
Qty Manufactured: 4,142 Kits

Abbott List Number: 02K47-25  
Abbott Commodity Number: 39-6271  
Expiration Date: 2017-09-30  
Package Insert No./Rev. G1-0392 / R04  
Test Date 2016/03/24

| Description    | Lot Number | Expiration Date |
|----------------|------------|-----------------|
| Microparticles | 26732UN16G | 2017-09-30      |
| Conjugate      | 26732UN16H | 2017-09-30      |
| ASD Reagent    | 26732UN16J | 2017-09-30      |

| Specification    | Specification Limit | Result                       |
|------------------|---------------------|------------------------------|
| Negative Control | 0.39 - 1.19 IU/mL   | 0.60 Pass                    |
| Negative Control | Lot Number          | 799343                       |
| Positive Control | 55 - 95 IU/mL       | 72 Pass                      |
| Positive Control | Lot Number          | 799342                       |
| Calibrator       | Lot Number          | 613178-613180, 613321-613323 |

For in-vitro diagnostic use.

Anti-TPO Reagents contain animal source material.

The lot meets all current release specifications per quality document MS.MASTER342.

The above product has been manufactured in compliance with Fisher Diagnostics' operating procedure.



Quality Representative

Date: 2016-03-28