


CERTIFICATE OF ANALYSIS
CEDIA METHAMPHETAMINE OFT MCC

CATALOG NO.	10014949
KIT LOT NO.	72046014
EXP DATE	2017 01

ITEM	LOT#	DOM	EXP DATE
ENZYME ACCEPTOR	72021529	12/15/14	2017 12
ENZYME DONOR	72045992	01/16/15	2017 01
EA RECONSTITUTION BUFFER	72046014	06/08/15	2017 01
ED RECONSTITUTION BUFFER	72046014	01/22/15	2017 01

Kit Testing

PARAMETER	SPECIFICATIONS	RESULT	PASS/FAIL
Low Control - Cutoff Cal	≥ 18 mA/min	59 mA/min	Pass
Cutoff Cal - High Control	≥ 18 mA/min	39 mA/min	Pass



QC SUPERVISOR06/22/15

Date