

Certificate of Analysis

Product Name: DTIBC Reagent 2
Product Numbers: 906706-155
Lot Number: 911967

Date of Manufacture: 10-OCT-16
Expiration Date: 2018-04-10
Storage: 2 - 8°C

Specification	Specification Limit	Result
Blank Absorbance @ 660 nm	<= 0.16 abs	0.06
Visual Clarity	Clear, colorless liquid free of particulates	Complies
Low Level Control Serum	Within Established Range	Complies
High Level Control Serum	Within Established Range	Complies
Post-Filling Bioburden	Acceptable	Complies
DTIBC Reagent 1	Lot Number	911966

For in-vitro diagnostic use.

The above lot number was manufactured in compliance with Quality Systems Regulations, has met all product specifications and has been released.

Storage Restriction: Do Not Freeze



Quality Representative

Date: 2016-10-21