

Tel 585-586-8800 Fax 585-899-7605 75 Panorama Creek Drive, Rochester, NY 14625

Product Certificate Thermo Scientific Nalgene and Nunc Products

Thermo Fisher Scientific hereby certifies that the product identified below is produced, inspected and found to be in compliance with product and quality specification requirements as documented in our ISO 13485:2003 Quality Management System (QMI-SAI Global File No. 1606319 and 1606321) in the USA.

Janet Coagroue

Janet Cosgrove
Raw Materials and
Product Compliance Manager

The following information represents Product Certification for: Item#: 322004-9050

Certificate issued: 03/02/2016

Description: BTL PKG SW AMB HDPE;1/2OZ,15ML Lot#: 1171008

Manufactured: 02/16/2016

| Part Number | Description | Common Name | DMF# | Cytotoxicity | USP Class VI | FDA Compliance - 21 CFR |
|-------------|-------------------------------|---------------------------|-------|--------------|--------------|---|
| 1-0435-99P | BTL,15ML,RND,N/M,HDPE,AMBER | COMPONENT PART | | | | |
| 8-0042-45P | RESIN,HDPE,AMBER,IBM,EBM,1.5% | COMPONENT PART | | | | |
| 8-0042-31 | RESIN,HDPE,IBM,EBM,EXT | HIGH-DENSITY POLYETHYLENE | 3310 | PASSED | PASSED | 177.1520 (c) 3.2a |
| 8-0097-17 | COLOR, AMBER, MULTI | COLORANT, OPAGUE, AMBER | 24299 | PASSED | PASSED | 177.1520, 177.1350,178.3297, 181.28 |
| 1-1811-62 | CLOS,20/415,PP,AMB,NALGE | COMPONENT PART | | | | |
| 8-0071-14P | RESIN, PP, AMB, INJ | COMPONENT PART | | | | |
| 8-0071-06 | Resin,PP,Inj | POLYPROPYLENE, INJECTION | 9988 | PASSED | PASSED | 177.1520(a)(1)(i), (c)1.1a,177.1520(b), (use conditionsA-H) |
| 8-0097-17 | COLOR,AMBER, MULTI | COLORANT, OPAGUE, AMBER | 24299 | PASSED | PASSED | 177.1520, 177.1350,178.3297, 181.28 |

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.