

Certificate of Analysis

Product Name: WBC Lyse
OEM Part Number: 08H521H01
OEM Lot Number: 57674UN16
Storage: 15 - 30°C
Qty Manufactured: 4,565 Kits

Abbott List Number: 08H52-01
Abbott Commodity Number: 39-6978
Expiration Date: 2017-11-30
Package Insert No./Rev. N/A / N/A

Specification	Specification Limit	Result
pH: Beg Sample-Value 1	8.5 - 8.7	8.6 Pass
pH: Beg Sample-Value 2	8.5 - 8.7	8.6 Pass
pH: End Sample-Value 1	8.5 - 8.7	8.6 Pass
pH: End Sample-Value 2	8.5 - 8.7	8.6 Pass
Osmolality: Beg Sample-Value 1	40 - 52 mOsmol/Kg	45 Pass
Osmolality: Beg Sample-Value 2	40 - 52 mOsmol/Kg	45 Pass
Osmolality: End Sample-Value 1	40 - 52 mOsmol/Kg	45 Pass
Osmolality: End Sample-Value 2	40 - 52 mOsmol/Kg	45 Pass
Conductivity: Beg Sample-Value 1	115 - 132 µmhos/cm	125 Pass
Conductivity: Beg Sample-Value 2	115 - 132 µmhos/cm	125 Pass
Conductivity: End Sample-Value 1	115 - 132 µmhos/cm	128 Pass
Conductivity: End Sample-Value 2	115 - 132 µmhos/cm	129 Pass
Date of Manufacture of Bulk	Report	2016-05-03
Test Date	Report	2016-05-05

1 Kit = 1 cubic

This product has been manufactured in compliance with Quality System Regulations, passed all Fisher Diagnostics final release specifications, and has met all Abbott acceptance requirements and current release specifications per applicable quality documents.

The product name, lot number, and expiration date matches the product name, lot number, and expiration date listed on this Certificate of Analysis.

Connie Carroll

Quality Representative

Date: 2016-05-06